



IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product Electronic measuring equipment

Name and address of the applicant E. ZOLLER GmbH & Co. KG

Gottlieb-Daimler-Str. 19 74385 Pleidelsheim

GERMANY

Name and address of the manufacturer E. ZOLLER GmbH & Co. KG

Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY

Name and address of the factory E. ZOLLER GmbH & Co. KG

Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY

Ratings and principal characteristics see next page

Trade mark (if any) ZOLLER

Customer's Testing Facility (CTF) Stage used CTF STAGE 1

Model/type Ref. genius, threadCheck

A sample of the product was tested and found

to be in conformity with

IEC 61010-1:2010

IEC 61010-1:2010/AMD1:2016

as shown in the Test Report Ref. No. which forms part of this certificate

713181356

Page 1 of 2


This CB Test Certificate is issued by the National Certification Body

CB 095184 0032 Rev. 00

Date, 2021-02-25









IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Ratings and principal characteristics (continued)

Rated voltage: 100-120 V and 200-240 V

Rated frequency: 50/60 Hz
Rated power: 600 VA
Protection class: I
Overvoltage category: II
Pollution degree: 2
Air pressure: 6-8 bar

Ambient: 5-30 °C, 5-90 % r.H (non-condensing)

Condition of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The detachable power cord set shall be complying with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard. The detachable power cord set shall not longer than 3 m.

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0032 Rev. 00

Date, 2021-02-25



